## Trigger logic for the $S\pi RIT$ experiments

C. Santamaria,<sup>\*1,\*2</sup> P. Pawlowski,<sup>\*3,\*2</sup> W. Lynch,<sup>\*1</sup> K. Pelczar,<sup>\*3</sup> J. Barney,<sup>\*1,\*2</sup> G. Cerizza,<sup>\*1,\*2</sup> J. Estee,<sup>\*1,\*2</sup>

T. Isobe,<sup>\*2</sup> M. Kaneko,<sup>\*4,\*2</sup> M. Kurata-Nishimura,<sup>\*2</sup> P. Lasko,<sup>\*3,\*2</sup> J. Lukasik,<sup>\*3</sup> A.B. McIntosh,<sup>\*5</sup>

T. Murakami,<sup>\*4,\*2</sup> M. B. Tsang,<sup>\*1</sup> S.J. Yennello,<sup>\*5,\*6</sup> Y. Zhang<sup>\*7,\*2</sup> for the SπRIT collaboration

The  $S\pi RIT$  experiments<sup>1)</sup> require a trigger to select central collisions in order to probe the nuclear equation of state at high densities using heavy ion collisions. Central collisions are studied via a Time Projection Chamber (TPC) surrounded by ancillary detectors. They are triggered by suitable multiplicities of charged particles in the Kyoto plastic scintillators located on the sides of the TPC and KATANA plastic scintillator paddles located downstream of the TPC. An extensive description of the KATANA array is provided in Ref. 3. The trigger was vetoed by the presence of a projectile-like fragment in the three KATANA veto scintillators, which are downstream of the TPC and centered about the beam trajectory, or if there is a hit in the active collimator that shadows the target frame from the incident beam.

The ultimate trigger decision was made by an FPGA in a "trigger box", shown in Fig. 1 in the shaded gray box. The internal trigger logic of this trigger box is not shown. It required a signal from the beam in a Start Counter scintillator located upstream, a multiplicity signal from the EASIROC module servicing the Kyoto multiplicity array or a multiplicity in the KATANA plastic that exceeds a threshold multiplicity selected by user. For regular data runs, the trigger condition required a multiplicity greater than four in the Kyoto array and the multiplicity in KATANA was ignored. Moreover, this trigger was vetoed in the trigger box if heavy fragments hit any of the three KATANA veto scintillators or if the beam hit the active collimator upstream. To obtain minimum bias events and to study the trigger bias of this event selection, the values of the Kyoto multiplicity, the KATANA trigger multiplicity, and the KATANA beam veto paddle thresholds were varied in separate minimum bias runs for further analysis.

To avoid a loss of track information in the TPC, a Fast Trigger decision was made. This Fast Trigger AND coincidence, shown in the figure, required the Start Counter, none of the three KATANA veto signals and vetoed by a BUSY signal. This opened the gating grid driver (GGD)<sup>2</sup>) via a TTL\_open signal generated by the latch G&D 2. The TTL\_open is normally terminated by G&D 1 after 11  $\mu$ sec but can be shortened

- \*6 Department of Chemistry, Texas A&M University
- \*7 Department of Physics, Tsinghua University

by the arrival of a random beam particle in one of the KATANA veto paddles or the lack of a trigger signal from the trigger box. The 11  $\mu$ sec maximum length of the gate is calculated from the 9  $\mu$ sec drift length of the TPC plus some safety margin to make sure all the volume of the TPC is included in the accepted time window. The TTL\_close signal, generated by the G&D 4, starts on the trailing edge of the TTL\_open signal and defines the closing of the GGD. The 5  $\mu$ sec width of the TTL\_close gate is calculated as the maximum drift time in the TPC for a heavy fragment to avoid these tracks from producing a signal in the next event.




Fig. 1. Schematic circuit diagram of the trigger logic used during the  $S\pi RIT$  experiments.

Three data acquisitions (DAQ) are coupled during the experiment to write data from the TPC (S $\pi$ RIT), the beam tracking counters and Kyoto array (RIKEN) and the KATANA array (KATANA). The BUSY signal is generated from an OR of the three DAQ BUSY signals plus the TTL\_open and TTL\_close. The trigger sent to the DAQ systems is a coincidence of the trigger from the trigger box, a NOT BUSY signal, and a delayed Start Counter signal in order to keep the Start Counter as the reference time for the DAQ.

During the experiment, the data acquisition successfully selected heavy ion collisions with on average a 800  $\mu$ s dead time. Analysis of data runs will quantitatively determine the average impact parameter of the heavy ion collisions in our setup and a further study of the minimum bias runs will constrain the bias of our experimental detection system.

References

- 1) R. Shane et al., Nucl. Instr. Meth. A 784, 513 (2015).
- S. Tangwancharoen et al., Nucl. Instr. Meth. A 853, 44 (2017).
- 3) P. Lasko et al., Nucl. Instr. Meth. A 856, 92 (2017).

<sup>\*1</sup> NSCL and Department of Physics and Astronomy, Michigan State University

<sup>\*&</sup>lt;sup>2</sup> RIKEN Nishina Center

<sup>\*&</sup>lt;sup>3</sup> Institute of Nuclear Physics, PAN

<sup>\*4</sup> Department of Physics, Kyoto University

<sup>\*5</sup> Cyclotron Institute, Texas A&M University